

A Comprehensive Surgical Imaging Experience

Access to View More...



Technical Specifications

| | |
|---------------------|---|
| Generator Frequency | 110kHz |
| Power Output | 3.6KW |
| Operating Modes | Continuous Fluoro, Boost Fluoro, Snap Mode, RAD Mode, Pulsed Fluoro |
| Image Receptor | Dynamic FPD |
| Size | 21cmX21cm |
| Imaging Chain | 1.5k X 1.5k |
| Image Memory | DFCC |
| Collimator | Parallel Shutter (Optional) |
| Laser Aimer | Yes |
| Power Supply | 230 VAC ± 10% :50Hz Frequency |

*Technical specification subject to change



APPROVED BY
 AERB & BIS

TECHNICAL SPECIFICATIONS DATA SHEET



Follow us @bplmedtech



www.bplmedicaltechnologies.com



Happier Living Everyday